

**Comparison of measured values of noise**

All values in mV

Channel1							
Cold		Warmed		Cold		Warmed	
700V				900V			
Vpp	8.31	Vpp	7.38	Vpp	4.13	Vpp	4.31
Vavg	0.04	Vavg	-0.75	Vavg	0.43	Vavg	0.40
Vstd	0.33	Vstd	0.29	Vstd	0.35	Vstd	0.36
Channel2							
Cold		Warmed		Cold		Warmed	
700V				900V			
Vpp	7.06	Vpp	9.00	Vpp	4.13	Vpp	5.06
Vavg	1.18	Vavg	0.40	Vavg	0.63	Vavg	1.64
Vstd	0.84	Vstd	0.70	Vstd	0.38	Vstd	0.53