

Date: 16,11,01

Time: 15:00 - 16:00 Temperování -> 21:00 konec

Crate: 01002

Absolute precision				Line regulation +/- 10%						Load regulation												
Nominal	Nominal +0.25%	Nominal -0.25%	%	Nominal +0.01%	Nominal -0.01%	-10%	%	+10%	%	19.5mA	18mA	11mA	4mA	2.1mA	0mA	Nominal +0.02%	dU4 [%]	Nominal -0.02%	dU18 [%]			
830,1515	832,0750	827,8250	0,0183 OK	830,2345	830,0685	830,1518	0,0000 OK	830,1519	0,0000 OK													
950,1725	952,3750	947,8250	0,0182 OK	950,2675	950,0775	950,1721	0,0000 OK	950,1730	0,0001 OK													
829,9826	832,0750	827,8250	-0,0021 OK	830,0656	829,8996	829,9825	0,0000 OK	829,9834	0,0001 OK													
949,9530	952,3750	947,8250	-0,0049 OK	950,0480	949,8580	949,9532	0,0000 OK	949,9550	0,0002 OK													
830,0467	832,0750	827,8250	0,0056 OK	830,1297	829,9637	830,0468	0,0000 OK	830,0467	0,0000 OK													
950,0579	952,3750	947,8250	0,0061 OK	950,1529	949,9629	950,0569	-0,0001 OK	950,0580	0,0001 OK													
830,0344	832,0750	827,8250	0,0041 OK	830,1174	829,9514	830,0339	-0,0001 OK	830,0349	0,0001 OK													
949,9838	952,3750	947,8250	-0,0017 OK	950,0788	949,8888	949,9846	0,0001 OK	949,9840	0,0000 OK													
830,1338	832,0750	827,8250	0,0161 OK	830,2168	830,0506	830,1337	0,0000 OK	830,1334	0,0000 OK													
950,1868	952,3750	947,8250	0,0197 OK	950,2619	950,0919	950,1868	0,0000 OK	950,1868	0,0000 OK													
830,0323	832,0750	827,8250	0,0039 OK	830,1153	829,9493	830,0327	0,0000 OK	830,0320	0,0000 OK													
950,0020	952,3750	947,8250	0,0002 OK	950,0970	949,9070	950,0030	0,0001 OK	950,0019	0,0000 OK													
830,0550	832,0750	827,8250	0,0068 OK	830,1380	829,9720	830,0549	0,0000 OK	830,0545	-0,0001 OK													
950,0549	952,3750	947,8250	0,0058 OK	950,1499	949,9599	950,0552	0,0000 OK	950,0545	0,0000 OK													
829,9879	832,0750	827,8250	-0,0003 OK	830,0809	829,9149	829,9893	0,0002 OK	829,9890	0,0000 OK	829,9445	829,9689	830,0830	830,1970	830,2280	830,3320	830,1639	0,014	829,8319	-0,014	OK		
949,9889	952,3750	947,8250	-0,0012 OK	950,0839	949,8939	949,9898	0,0000 OK	949,9891	0,0000 OK	949,9423	949,9664	950,0789	950,1915	950,2220	950,3250	950,1789	0,012	949,7989	-0,012	OK		
830,0615	832,0750	827,8250	0,0074 OK	830,1445	829,9785	830,0618	0,0000 OK	830,0611	0,0000 OK													
950,0769	952,3750	947,8250	0,0051 OK	950,1719	949,9819	950,0755	-0,0001 OK	950,0770	0,0000 OK													
830,0212	832,0750	827,8250	0,0029 OK	830,1042	829,9382	830,0214	0,0000 OK	830,0215	0,0000 OK													
950,0161	952,3750	947,8250	0,0017 OK	950,1111	949,9211	950,0151	-0,0001 OK	950,0175	0,0001 OK													
830,0720	832,0750	827,8250	0,0087 OK	830,1530	829,9860	830,0725	0,0001 OK	830,0729	0,0001 OK													
950,0743	952,3750	947,8250	0,0078 OK	950,1893	949,9793	950,0742	0,0000 OK	950,0741	0,0000 OK													
829,9266	832,0750	827,8250	-0,0088 OK	830,0099	829,8439	829,9263	-0,0001 OK	829,9270	0,0000 OK													
949,9141	952,3750	947,8250	-0,0090 OK	950,0091	949,8191	949,9140	0,0000 OK	949,9420	0,0029 OK													
830,0552	832,0750	827,8250	0,0087 OK	830,1382	829,9722	830,0555	0,0000 OK	830,0550	0,0000 OK													
950,0582	952,3750	947,8250	0,0059 OK	950,1512	949,9612	950,0584	0,0000 OK	950,0588	0,0000 OK													
830,0047	832,0750	827,8250	0,0006 OK	830,0877	829,9217	830,0052	0,0001 OK	830,0049	0,0000 OK													
949,9820	952,3750	947,8250	-0,0019 OK	950,0770	949,8870	949,9825	0,0001 OK	949,9812	-0,0001 OK													
830,0717	832,0750	827,8250	0,0086 OK	830,1547	829,9887	830,0722	0,0001 OK	830,0712	-0,0001 OK													
950,0780	952,3750	947,8250	0,0082 OK	950,1730	949,9830	950,0800	0,0002 OK	950,0771	-0,0001 OK													
830,0248	832,0750	827,8250	0,0030 OK	830,1078	829,9418	830,0248	0,0000 OK	830,0257	0,0001 OK	829,9923	830,0153	830,1225	830,2298	830,2589	830,3675	830,1908	0,013	829,8586	-0,013	OK		
950,0330	952,3750	947,8250	0,0035 OK	950,1280	949,9380	950,0328	0,0000 OK	950,0340	0,0001 OK	949,9991	950,0219	950,1280	950,2342	950,2630	950,3660	950,2230	0,011	949,8430	-0,011	OK		