

Date: 17.11.01

Time: 08:00 - 09:00, 09:00 - 12:00

Temperování -> Měření

Crate: 01009

Absolute precision				Line regulation +/- 10%						Load regulation													
Nominal	Nominal +0.25%	Nominal -0.25%	%	Nominal +0.01%	Nominal -0.01%	-10%	%	+10%	%	19.5mA	18mA	11mA	4mA	2.1mA	0mA	Nominal +0.02%	dU4 [%]	Nominal -0.02%	dU18 [%]				
830.0520	832.0750	827.9250	0.0099	OK	830.1650	829.9990	830.0818	0.0000	OK	830.0816	0.0000	OK											
950.0775	952.3750	947.6250	0.0082	OK	950.1725	949.9825	950.0773	0.0000	OK	950.0768	-0.0001	OK											
830.0895	832.0750	827.9250	0.0109	OK	830.1725	830.0065	830.0903	0.0001	OK	830.0900	0.0001	OK											
950.0808	952.3750	947.6250	0.0085	OK	950.1758	949.9858	950.0805	0.0000	OK	950.0810	0.0000	OK											
830.0274	832.0750	827.9250	0.0033	OK	830.1104	829.9444	830.0273	0.0000	OK	830.0272	0.0000	OK											
950.0145	952.3750	947.6250	0.0015	OK	950.1095	949.9195	950.0142	0.0000	OK	950.0153	0.0001	OK											
830.0697	832.0750	827.9250	0.0084	OK	830.1527	829.9867	830.0698	0.0000	OK	830.0698	0.0000	OK											
950.0442	952.3750	947.6250	0.0047	OK	950.1392	949.9492	950.0439	0.0000	OK	950.0441	0.0000	OK											
830.1260	832.0750	827.9250	0.0152	OK	830.2060	830.0430	830.1255	-0.0001	OK	830.1264	0.0000	OK											
950.1558	952.3750	947.6250	0.0164	OK	950.2508	950.0808	950.1565	0.0001	OK	950.1555	0.0000	OK											
830.0440	832.0750	827.9250	0.0053	OK	830.1270	829.9610	830.0445	0.0001	OK	830.0436	0.0000	OK											
950.0364	952.3750	947.6250	0.0038	OK	950.1314	949.9414	950.0359	-0.0001	OK	950.0358	-0.0001	OK											
830.0918	832.0750	827.9250	0.0111	OK	830.1748	830.0088	830.0920	0.0000	OK	830.0916	0.0000	OK											
950.0999	952.3750	947.6250	0.0105	OK	950.1949	950.0049	950.0994	-0.0001	OK	950.1005	0.0001	OK											
830.0732	832.0750	827.9250	0.0088	OK	830.1562	829.9902	830.0735	0.0000	OK	830.0729	0.0000	OK											
950.0497	952.3750	947.6250	0.0052	OK	950.1447	949.9547	950.0495	0.0000	OK	950.0500	0.0000	OK											
830.0212	832.0750	827.9250	0.0026	OK	830.1042	829.9382	830.0210	0.0000	OK	830.0218	0.0001	OK											
950.0154	952.3750	947.6250	0.0016	OK	950.1104	949.9204	950.0153	0.0000	OK	950.0156	0.0000	OK											
830.0508	832.0750	827.9250	0.0061	OK	830.1338	829.9678	830.0510	0.0000	OK	830.0515	0.0001	OK											
950.0450	952.3750	947.6250	0.0047	OK	950.1400	949.9500	950.0445	-0.0001	OK	950.0455	-0.0001	OK											
830.0744	832.0750	827.9250	0.0090	OK	830.1574	829.9914	830.0740	0.0000	OK	830.0739	-0.0001	OK											
950.0747	952.3750	947.6250	0.0079	OK	950.1697	949.9797	950.0749	0.0000	OK	950.0755	0.0001	OK											
829.9832	832.0750	827.9250	-0.0020	OK	830.0662	829.9002	829.9836	0.0000	OK	829.9830	0.0000	OK											
949.9849	952.3750	947.6250	-0.0016	OK	950.0799	949.8999	949.9843	-0.0001	OK	949.9853	0.0000	OK											
830.0918	832.0750	827.9250	0.0111	OK	830.1748	830.0088	830.0922	0.0000	OK	830.0916	0.0000	OK											
950.0997	952.3750	947.6250	0.0105	OK	950.1947	950.0047	950.0996	0.0000	OK	950.0998	0.0000	OK											
830.0136	832.0750	827.9250	0.0016	OK	830.0966	829.9306	830.0148	0.0001	OK	830.0140	0.0000	OK											
949.9908	952.3750	947.6250	-0.0010	OK	950.0858	949.8958	949.9906	0.0000	OK	949.9910	0.0000	OK											
830.0762	832.0750	827.9250	0.0092	OK	830.1592	829.9932	830.0769	0.0001	OK	830.0760	0.0000	OK											
950.0930	952.3750	947.6250	0.0098	OK	950.1880	949.9980	950.0923	-0.0001	OK	950.0945	0.0002	OK											
830.0859	832.0750	827.9250	0.0103	OK	830.1689	830.0029	830.0860	0.0000	OK	830.0855	0.0000	OK	830.0399	830.0642	830.1776	830.2910	830.3218	830.4151	830.2519	0.014	829.9199	-0.014	OK
950.0785	952.3750	947.6250	0.0083	OK	950.1735	949.9835	950.0800	0.0002	OK	950.0799	0.0001	OK	950.0331	950.0569	950.1679	950.2789	950.3090	950.4080	950.2885	0.012	949.8885	-0.012	OK