

Date: 19.11.01

Time: 08:00 - 09:30 11:00

Temperování

Měření

Crate: 01013

Nominal	Absolute precision			Line regulation +/- 10%						Load regulation									
	Nominal +0.25%	Nominal -0.25%	%	Nominal +0.01%	Nominal -0.01%	-10%	%	+10%	%	19.5mA	18mA	11mA	4mA	2.1mA	0mA	Nominal +0.02%	HU4 [%]	Nominal -0.02%	dU18 [%]
829.9975	832.0750	827.9250	-0.0003 OK	830.0805	829.9145	829.9971	0.0000 OK	829.9969	-0.0001 OK	820.0000	820.0000	820.0000	820.0000	820.0000	820.0000	830.1635	0.000	829.8315	0.000 OK
950.0350	952.3750	947.6250	0.0037 OK	950.1300	949.9400	950.0353	0.0000 OK	950.0350	0.0000 OK	940.0000	940.0000	940.0000	940.0000	940.0000	940.0000	950.2250	0.000	949.8450	0.000 OK
830.0442	832.0750	827.9250	0.0053 OK	830.1272	829.9512	830.0435	-0.0001 OK	830.0437	-0.0001 OK	820.0000	820.0000	820.0000	820.0000	820.0000	820.0000	830.2102	0.000	829.8782	0.000 OK
949.9558	952.3750	947.6250	-0.0004 OK	950.0908	949.9008	949.9555	0.0000 OK	949.9562	0.0000 OK	940.0000	940.0000	940.0000	940.0000	940.0000	940.0000	950.1858	0.000	949.8058	0.000 OK
830.0068	832.0750	827.9250	0.0012 OK	830.0928	829.9288	830.0096	0.0000 OK	830.0099	0.0000 OK	820.0000	820.0000	820.0000	820.0000	820.0000	820.0000	830.1758	0.000	829.8438	0.000 OK
950.0247	952.3750	947.6250	0.0026 OK	950.1197	949.9297	950.0237	-0.0001 OK	950.0245	0.0000 OK	940.0000	940.0000	940.0000	940.0000	940.0000	940.0000	950.2147	0.000	949.8347	0.000 OK
829.9780	832.0750	827.9250	-0.0027 OK	830.0610	829.8950	829.9777	0.0000 OK	829.9784	0.0000 OK	820.0000	820.0000	820.0000	820.0000	820.0000	820.0000	830.1440	0.000	829.8120	0.000 OK
949.9686	952.3750	947.6250	-0.0033 OK	950.0636	949.8736	949.9683	0.0000 OK	949.9680	-0.0001 OK	940.0000	940.0000	940.0000	940.0000	940.0000	940.0000	950.1586	0.000	949.7786	0.000 OK
830.1218	832.0750	827.9250	0.0147 OK	830.2048	830.0388	830.1220	0.0000 OK	830.1160	-0.0007 OK	820.0000	820.0000	820.0000	820.0000	820.0000	820.0000	830.2878	0.000	829.9558	0.000 OK
950.1332	952.3750	947.6250	0.0140 OK	950.2282	950.0382	950.1339	0.0001 OK	950.1330	0.0000 OK	940.0000	940.0000	940.0000	940.0000	940.0000	940.0000	950.3232	0.000	949.9432	0.000 OK
830.0693	832.0750	827.9250	0.0083 OK	830.1523	829.9863	830.0703	0.0001 OK	830.0691	0.0000 OK	820.0000	820.0000	820.0000	820.0000	820.0000	820.0000	830.2353	0.000	829.9033	0.000 OK
950.0562	952.3750	947.6250	0.0059 OK	950.1512	949.9612	950.0558	0.0000 OK	950.0564	0.0000 OK	940.0000	940.0000	940.0000	940.0000	940.0000	940.0000	950.2462	0.000	949.8662	0.000 OK
829.8861	832.0750	827.9250	-0.0137 OK	829.9691	829.8031	829.8858	0.0000 OK	829.8864	0.0000 OK	820.0000	820.0000	820.0000	820.0000	820.0000	820.0000	830.0521	0.000	829.7201	0.000 OK
949.8679	952.3750	947.6250	-0.0139 OK	949.9629	949.7729	949.8680	0.0000 OK	949.8675	0.0000 OK	940.0000	940.0000	940.0000	940.0000	940.0000	940.0000	950.0579	0.000	949.6779	0.000 OK
829.9475	832.0750	827.9250	-0.0063 OK	830.0305	829.8645	829.9471	0.0000 OK	829.9473	0.0000 OK	820.0000	820.0000	820.0000	820.0000	820.0000	820.0000	830.1135	0.000	829.7815	0.000 OK
949.9193	952.3750	947.6250	-0.0065 OK	950.0143	949.8243	949.9189	0.0000 OK	949.9200	0.0001 OK	940.0000	940.0000	940.0000	940.0000	940.0000	940.0000	950.1093	0.000	949.7293	0.000 OK
829.8577	832.0750	827.9250	-0.0171 OK	829.9407	829.7747	829.8586	0.0001 OK	829.8570	-0.0001 OK	820.0000	820.0000	820.0000	820.0000	820.0000	820.0000	830.0237	0.000	829.6917	0.000 OK
949.8281	952.3750	947.6250	-0.0181 OK	949.9231	949.7331	949.8277	0.0000 OK	949.8285	0.0000 OK	940.0000	940.0000	940.0000	940.0000	940.0000	940.0000	950.0181	0.000	949.6381	0.000 OK
829.9733	832.0750	827.9250	-0.0032 OK	830.0563	829.8903	829.9730	0.0000 OK	829.9726	-0.0001 OK	820.0000	820.0000	820.0000	820.0000	820.0000	820.0000	830.1393	0.000	829.8073	0.000 OK
949.9724	952.3750	947.6250	-0.0029 OK	950.0674	949.8774	949.9721	0.0000 OK	949.9708	-0.0002 OK	940.0000	940.0000	940.0000	940.0000	940.0000	940.0000	950.1624	0.000	949.7824	0.000 OK
830.0293	832.0750	827.9250	0.0035 OK	830.1123	829.9463	830.0291	0.0000 OK	830.0289	0.0000 OK	820.0000	820.0000	820.0000	820.0000	820.0000	820.0000	830.1953	0.000	829.8633	0.000 OK
950.0297	952.3750	947.6250	0.0031 OK	950.1247	949.9347	950.0291	-0.0001 OK	950.0300	0.0000 OK	940.0000	940.0000	940.0000	940.0000	940.0000	940.0000	950.2197	0.000	949.8397	0.000 OK
829.9249	832.0750	827.9250	-0.0090 OK	830.0079	829.8419	829.9244	-0.0603 BAD	829.9238	-0.0001 OK	820.0000	820.0000	820.0000	820.0000	820.0000	820.0000	830.0909	0.000	829.7589	0.000 OK
949.8827	952.3750	947.6250	-0.0123 OK	949.9777	949.7877	949.8823	0.0000 OK	949.8820	-0.0001 OK	940.0000	940.0000	940.0000	940.0000	940.0000	940.0000	950.0727	0.000	949.6927	0.000 OK
829.8465	832.0750	827.9250	-0.0185 OK	829.9295	829.7635	829.8467	0.0000 OK	829.8472	0.0001 OK	820.0000	820.0000	820.0000	820.0000	820.0000	820.0000	830.0125	0.000	829.6805	0.000 OK
949.8561	952.3750	947.6250	-0.0151 OK	949.9511	949.7611	949.8555	-0.0001 OK	949.8558	0.0000 OK	940.0000	940.0000	940.0000	940.0000	940.0000	940.0000	950.0461	0.000	949.6661	0.000 OK
829.8795	832.0750	827.9250	-0.0145 OK	829.9625	829.7965	829.8799	0.0000 OK	829.8793	0.0000 OK	820.0000	820.0000	820.0000	820.0000	820.0000	820.0000	830.0455	0.000	829.7135	0.000 OK
949.8362	952.3750	947.6250	-0.0172 OK	949.9312	949.7412	949.8365	0.0000 OK	949.8359	0.0000 OK	940.0000	940.0000	940.0000	940.0000	940.0000	940.0000	950.0262	0.000	949.6462	0.000 OK
830.1089	832.0750	827.9250	0.0131 OK	830.1919	830.0259	830.1095	0.0001 OK	830.1095	0.0001 OK	820.0000	820.0000	820.0000	820.0000	820.0000	820.0000	830.2749	0.000	829.9429	0.000 OK
950.1263	952.3750	947.6250	0.0133 OK	950.2213	950.0313	950.1255	-0.0001 OK	950.1262	0.0000 OK	940.0000	940.0000	940.0000	940.0000	940.0000	940.0000	950.3163	0.000	949.9363	0.000 OK
830.0028	832.0750	827.9250	0.0003 OK	830.0858	829.9198	830.0023	-0.0001 OK	830.0030	0.0000 OK	830.0385	830.0590	830.1544	830.2499	830.2758	830.3655	830.1688	0.011	829.8368	-0.011 OK
949.9941	952.3750	947.6250	-0.0008 OK	950.0891	949.8991	949.9922	0.0001 OK	949.9937	0.0000 OK	950.0350	950.0550	950.1483	950.2417	950.2870	950.3570	950.1841	0.010	949.8041	-0.010 OK